

CERTIFICATION

**LANGER**  
EMV-Technik

**Nöthnitzer Hang 31**  
**Germany 01728 Bannewitz**  
 Phone: +49 (351) 43 00 93 - 0  
 Fax: +49 (351) 43 00 93 - 22  
 email: mail@langer-emv.de

Cert. No. 090210QSHH150

Muster

This Probe has been individually calibrated using the following procedure for monitoring the frequency response:

**Publication:**  
Magnetic Near Field Probe

**Calibration**  
**Procedure:** LACPICR01  
**Uncertainty:** +/- 1.5 dB

**Date of Cal.:** Feb 02, 2012  
**Cal. Interval:** 12 months

**Model:** ICR HH150  
**Serial:** #9999999

**Manufacturer:**  
Langer EMV-Technik

**Temperature:** 20°C  
**Humidity:** 51 %

**Test and Measurement Equipment**

Manufacturer	Model	Serial	Due Date
Robde & Schwarz	FLS-6	100407	2010/03/30

TG = 100dBµV, Att = 5dB, RBW = 1MHz, VBW = 100kHz, SWT = 120ms, normalized

.....  
Dipl.-Ing. C. Stange  
Development

**LANGER**  
EMV-Technik GmbH  
Nöthnitzer Hang 31  
DE-01728 Bannewitz

This calibration is in compliance with the International EMC Standard of IECs IEC 61967. Supporting Documentation relative to traceability is on file and available for examination upon request.

This certificate shall not be reproduced except in full without the written approval of Langer EMV-Technik.

LANGER EMV-Technik GmbH / www.langer-emv.com IC Test System - IC-Scanner\_ICR Sonden 0313